

Air-Sea Interaction Simulation Facility for Research, Teaching and Instrument Development

Mark Donelan

Rosenstiel School of Marine and Atmospheric Science

University of Miami

4600 Rickenbacker Causeway

Miami, FL 33149-1098

phone: (305) 361-4717 fax: (305) 361-4701 e-mail: mdonelan@rsmas.miami.edu

Hans C. Graber

Rosenstiel School of Marine and Atmospheric Science

University of Miami

4600 Rickenbacker Causeway

Miami, FL 33149-1098

phone: (305) 361-4935 fax: (305) 361-4701 e-mail: hgraber@rsmas.miami.edu

Eric S. Saltzman

Rosenstiel School of Marine and Atmospheric Science

University of Miami

4600 Rickenbacker Causeway

Miami, FL 33149-1098

phone: (305) 361-4922 fax: (305) 361-4689 e-mail: esaltzman@rsmas.miami.edu

Otis B. Brown

Rosenstiel School of Marine and Atmospheric Science

University of Miami

4600 Rickenbacker Causeway

Miami, FL 33149-1098

phone: (305) 361-4000 fax: (305) 361-4711 e-mail: obrown@rsmas.miami.edu

Award # N00014-98-1-0261

LONG-TERM GOAL

To provide a modern air-sea interaction simulation facility with salt water and high wind speed capability to: enable a wide range of research in remote sensing, gas transfer, surface chemistry, spray and aerosol generation, turbulence and wave dynamics; to enhance and enliven the teaching of processes at and near the air-water interface and the technology and methodology of measurement and analysis of thereof; to provide a versatile test-bed for the development, testing and calibration of instruments.

APPROACH

We propose to build a medium size wind-water tunnel with a working section of 1 m x 1 m x 15 m (Figure 1). The school has allocated the necessary space for the facility (overall length 24 m) and associated space for physics, chemistry and computing laboratories. The tunnel/tank and associated

Report Documentation Page				Form Approved OMB No. 0704-0188	
Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.					
1. REPORT DATE 1998		2. REPORT TYPE		3. DATES COVERED 00-00-1998 to 00-00-1998	
4. TITLE AND SUBTITLE Air-Sea Interaction Simulation Facility for Research, Teaching and Instrument Development				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S)				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) University of Miami,Rosenstiel School of Marine and Atmospheric Science,4600 Rickenbacker Causeway,Miami,FL,33149				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited					
13. SUPPLEMENTARY NOTES See also ADM002252.					
14. ABSTRACT					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT Same as Report (SAR)	18. NUMBER OF PAGES 4	19a. NAME OF RESPONSIBLE PERSON
a. REPORT unclassified	b. ABSTRACT unclassified	c. THIS PAGE unclassified			

return ducting and piping will be made of acrylic or stainless steel to minimize corrosion and chemical contamination of trace species. The facility is designed to use sea water or fresh water as desired and the location chosen is already supplied with storage tanks and pumping systems for both sea water and fresh water.




Figure 1. Schematic drawing of the air-sea interaction, salt-water tank (ASIST) facility.

A large heat pump will be provided to allow us to maintain water temperatures between 0°C and 40 °C, which will allow us to explore temperature effects on the physical and chemical aspects of gas transfer. The air and water flows will be driven by a 40 hp variable fan and pump respectively. This will allow tests to be done at high wind speeds (30 m/s at centerline of the tunnel -- equivalent to hurricane speeds) and also substantial water speeds (0 to 35 cm/s) permitting tests of the effect of water turbulence on gas transfer.

The wind tunnel is designed to operate in two modes: 'recirculating' and 'venting'. This flexibility makes it possible to do both recirculation experiments of evasion and invasion, in which the equilibration of a closed system provides an integrated measure of gas transfer -- leading to great precision, or venting experiments (generally for evasion), in which a large concentration difference can be maintained for extended periods.

The return air duct is displaced laterally rather than directly above the working section. This will permit the maximum access to viewing the surface from above for remote sensing experiments and to introducing profiling *in situ* probes. Each 2.5 m segment of the working section is to be constructed in two parts: 1) the bottom and sides are to be of a single piece; 2) the top will be a removable slab. This offers maximum access to the tank for cleaning and allows complete flexibility in the choice of water depth between 0 and 1 m. The return air will run outside the building in a well insulated duct. This both optimizes the use of valuable inner building space and simplifies the periodic duct cleaning.

The Air-Sea Interaction Facility will have the following characteristics:

1. acrylic and stainless steel construction.
2. sea water or fresh water and all mixtures.
3. controlled water temperature 0 °C to 40 °C.
4. wind speeds of 0 - 30 m/s.
5. water speeds of 0 - 35 cm/s in either direction.
6. programmable wave-maker -- hydraulically driven paddle.
7. recirculating or venting modes of operation.
8. attached chemistry laboratory, instrument laboratory and data collecting and computing room.

WORK COMPLETED

The wind-wave tank facility has been designed (see Figure 1) and several components have been ordered. The area of the ground floor of the Glassell building, which has been allocated for the installation of the facility, has been cleared of its previous occupants and the walls and ceiling have been cleaned and repainted. The installation will begin in late November, while the purchase of measurement and data acquisition systems will begin in December.

RESULTS

None yet.

IMPACT/APPLICATION

This salt-water capable air-sea interaction facility will be used for research, teaching, and for instrument testing, calibration, and development. The facility will be unique in its capacity to simulate air-sea interaction over a wide range of environmental conditions using both seawater and fresh water. This facility will be used by a number of ongoing research programs investigating the remote sensing of surface ocean wind/wave fields, turbulent kinetic energy dissipation in near surface waters, buoy technology development for high resolution wave and flux measurements, skin temperature effects on remote sensing, and air-sea exchange of gases and aerosols. The facility will also be incorporated into educational programs at the University of Miami in applied marine physics, physical oceanography, and marine and atmospheric chemistry.

TRANSITIONS

None yet.

RELATED PROJECTS

The air-sea interaction, salt-water tank (ASIST) facility will be a necessary adjunct to many field programs such as the Shoaling Waves Experiment (SHOWEX). The ability to control the external forcing to interfacial processes will be of enormous help in elucidating the role of such processes in air-sea interaction and remote sensing.